

12025 NE Marx St. Portland, OR 97220 503-253-3511 / www.greenleaflabs.com

License#: 10029074C70

Ugli Tea - Night Cap CBD

Sample ID: G1J0198-05

Matrix: Hemp Products

Test ID: 5015032 **Source ID:** UTN-3

Date Sampled: 10/08/21

Date Accepted: 10/08/21

Regulatory Compliance Testing (OAR-333-007) Official Report

Tropiceel Products Inc.

	Primary and Du	plicate Result	s at a Glance	
	Averaged	Primary	Duplicate	%RPD (15% Action Level)
Total THC:	<loq(0.0382%) %<="" td=""><td>< LOQ %</td><td>< LOQ %</td><td>0 % PASS</td></loq(0.0382%)>	< LOQ %	< LOQ %	0 % PASS
Total CBD:	0.7591 %	0.7746 %	0.7435 %	4.1 % PASS
Total CBG:	<loq(0.0346%) %<="" td=""><td>< LOQ %</td><td>< LOQ %</td><td></td></loq(0.0346%)>	< LOQ %	< LOQ %	





12025 NE Marx St. Portland, OR 97220 503-253-3511 / www.greenleaflabs.com

(OAR-333-007) **Official Report** License#: 10029074C70

Ugli Tea - Night Cap CBD

Sample ID: G1J0198-05

Matrix: Hemp Products


Test ID: 5015032 Source ID: UTN-3

Date Sampled: 10/08/21

Date Accepted: 10/08/21

Tropiceel Products Inc.

Regulatory Compliance Testing



Total THC = delta 9-THC + (THCA * 0.877) Total CBD = CBD + (CBDA * 0.877)

Total CBG = CBG + (CBGA * 0.878)

LOQ=Limit of Quantification, the lowest measurable concentration of an analyte.





Regulatory Compliance Testing (OAR-333-007) Official Report

503-253-3511 / www.greenleaflabs.com License#: 10029074C70

Quality Control Potency

Batch: 2142007 - 215-Products

Blank(2142007-BLK1)							
Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.0005	%		10/12/21 09:08	10/12/21 16:46	
delta 9-THC	< LOQ	0.0005	%		10/12/21 09:08	10/12/21 16:46	
THCV	< LOQ	0.0150	%		10/12/21 09:08	10/12/21 16:46	
THCVA	< LOQ	0.0224	%		10/12/21 09:08	10/12/21 16:46	
CBD	< LOQ	0.0005	%		10/12/21 09:08	10/12/21 16:46	
CBDA	< LOQ	0.0005	%		10/12/21 09:08	10/12/21 16:46	
CBDV	< LOQ	0.0154	%		10/12/21 09:08	10/12/21 16:46	
CBDVA	< LOQ	0.0212	%		10/12/21 09:08	10/12/21 16:46	
CBN	< LOQ	0.0138	%		10/12/21 09:08	10/12/21 16:46	
CBG	< LOQ	0.0161	%		10/12/21 09:08	10/12/21 16:46	
CBGA	< LOQ	0.0213	%		10/12/21 09:08	10/12/21 16:46	
CBC	< LOQ	0.0202	%		10/12/21 09:08	10/12/21 16:46	

Reference(2142007-SRM1)							
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
delta 9-THC	97.0	0.0004	%	80-120	10/12/21 09:08	10/12/21 17:09	
CBD	105	0.0004	%	80-120	10/12/21 09:08	10/12/21 17:09	







Regulatory Compliance Testing (OAR-333-007)
Official Report

License#: 10029074C70

manually for all samples.

Ammended Report Explanation

Sample name updated from "Ugli Tea - Blood Sugar CBG" to "Ugli Tea - Night Cap CBD" - RC 1.7.2022

Notes and Definitions

Regulatory Compliance samples were collected onsite at facility according to ORELAP-SOP-001 and ORELAP-SOP-002 and following Sampling Plan FN117.

Quality Control samples were tested as received.

ATM	Non-cannabis matrix related interference or suppression of Internal standard
BLI	Baseline Interference - Cannabinoid peak interference in chromatographic baseline affecting QC recovery
BLK	Analyte detected in method blank, but not associated samples.
BSH	Blank Spike High - Blank Spike recovery above method limit. no detections in samples.
BSL	Blank Spike Low - Blank Spike recovery below lower method limit, analyte chromatography reviewed
С	manually for all samples.
CBD	Interference due to co-elution
CV1	CBD matrix interference on GC Pest chromatography
CV2	CCV was above acceptance criteria, Non-detect samples are considered acceptable.
INF	CCV was below acceptance criteria, sample still exceeds regulatory limit.
ISH	One or more QC falls outside acceptance criteria. Data entered into LIMS for informational purposes only.
ISL	Internal Standard concentration is above acceptance criteria.
MSH	Internal Standard concentration is below acceptance criteria.
MSI	Matrix Spike High - Matrix Spike recovery above method limits.
MSL	Matrix Spike Interference - Matrix spike source sample contains analyte hit above calibration affecting
TPP	recovery accuracy in Matrix Spike.
U	Matrix Spike Low - Matrix Spike recovery below lower method limit, analyte chromatography reviewed

Internal Standard concentration outside control limit due to matrix interference

